



HAL
open science

High Oxygen Consumption in SARS-COV2: Could the Development of Low-Cost Oxygen Rebreather Be Considered?

Nicolas Vallée, Sarah Rives, Anne-Virginie Desruelle, Sebastian Marzetti, Valentin Barchasz, Jean-Jacques Risso, Valentin Gies

► To cite this version:

Nicolas Vallée, Sarah Rives, Anne-Virginie Desruelle, Sebastian Marzetti, Valentin Barchasz, et al.. High Oxygen Consumption in SARS-COV2: Could the Development of Low-Cost Oxygen Rebreather Be Considered?. *Frontiers in Physiology*, 2021, 11, 10.3389/fphys.2020.607913 . hal-03502819

HAL Id: hal-03502819

<https://univ-tln.hal.science/hal-03502819v1>

Submitted on 26 Dec 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



High Oxygen Consumption in SARS-COV2: Could the Development of Low-Cost Oxygen Rebreather Be Considered?

Nicolas Vallée^{1*}, Sarah Rives¹, Anne-Virginie Desruelle¹, Sebastian Marzetti², Valentin Barchasz², Jean-Jacques Risso¹ and Valentin Gies²

¹ Institut de Recherche Biomédicale des Armées_Equipe de Recherche Subaquatique Opérationnelle, Toulon, France,

² Université de Toulon, IM2NP, La Garde, France

Keywords: rebreather, ventilation, crisis, dioxygen, ICU

OPEN ACCESS

Edited by:

Giuseppe Andrea Miserocchi,
University of Milano-Bicocca, Italy

Reviewed by:

Carmen Silvia Valente Barbas,
University of São Paulo, Brazil

*Correspondence:

Nicolas Vallée
nicolas.vallee5@hotmail.fr

Specialty section:

This article was submitted to
Respiratory Physiology,
a section of the journal
Frontiers in Physiology

Received: 22 September 2020

Accepted: 21 December 2020


Published: 21 January 2021

Citation:

Vallée N, Rives S, Desruelle A-V,
Marzetti S, Barchasz V, Risso J-J and
Gies V (2021) High Oxygen
Consumption in SARS-COV2: Could
the Development of Low-Cost Oxygen
Rebreather Be Considered?
Front. Physiol. 11:607913.
doi: 10.3389/fphys.2020.607913

Oxygen is a first-line drug in the therapeutic management of patients with SARS-Cov2 (Tobin, 2020). Being limited, oxygen is a resource to save (Maves et al., 2020). In the same time, it should be remembered that medical care can become financially toxic (See and Slonim, 2020) because the cost of healthcare steadily rises (Mycyk, 2020). Is ventilators by lottery (Silva, 2020) the solution? In this way, requiring a low-cost respirator <200€ is attractive (Gies et al., 2020), but its value may be limited in view of the consumption, supply and cost of oxygen. The Swiss newspaper Le Temps reported in its article of July 15, 2020, whose title could be translated as “With Covid-19, the world faces an oxygen shortage,” that the oxygen demand in Italy has increased 10-fold or 14-fold and the costs are likely to soar (Logean, 2020).

We wonder viewing the **High oxygen consumption in SARS-COV2 if the development of low-cost oxygen rebreather could be considered?** The oxygen rebreather, or closed-circuit rebreather, is a simple apparatus well-known to the world of scuba diving and therefore the main goal is to be able to evolve a long time underwater with a small carry of oxygen (Wingelaar et al., 2017). The device is designed so that the majority of the oxygen delivered to the diver is consumed by the organism: the exhaled oxygen must be able to be re-inhaled through a counterlung, while the exhaled carbon dioxide must be able to be extracted using soda lime canister. When the counterlung is empty, the low pressure triggered by the diver’s inspiration is enough to activate a regulator connected to a pressurized oxygen cylinder, which replenishes the bladder instantly. Like respirators, one can imagine that a similar device could be added to it in order to limit the loss of oxygen (**Figure 1**). Since no gas exhaled by the patient is released into the atmosphere, this type of device can help improve the safety of healthcare workers. The development of such a device is simple but it must obviously consider the health safety rules with the use of filters against the virus, but also the rules in force in the field of diving. For example, it is important to properly calibrate the regulator in order to avoid any additional ventilatory effort, which could lead to alveolar damage (Wilmshurst, 2019): electronic systems can easily overcome the defects of the mechanical systems. It is also crucial to have the ability to change soda lime before it is saturated (Arieli, 2008): lime with colored indicators exists. The reaction of soda lime with CO₂ is exothermic (Shaw and Scott, 1998; Silvanus et al., 2019), but the length of the breathing circuits must be sufficient to dissipate



the heat. An ice bath around the soda lime canister could also be suitable. It must be taken into account that the exhaled gases are saturated with water, and that water traps for condensation may be necessary. Additionally, if we are looking to obtain a very high FiO_2 , it is also important to take into account the denitrogenation of the organism (Katz et al., 2017): the nitrogen regularly exhaled can be evacuated by means of regular flushing of the counterlung in the 1st h of treatment. Finally, most of intubated and ventilated COVID patients need levels of positive end-expiratory pressure (PEEP) around 10–15 cm H_2O because of their acute respiratory syndrome (Navas-Blanco and Dudaryk, 2020), proportional forces can be applied on the counter lung and the membrane regulator in order to achieve light pressures inside the circuit. These forces can take the appearance of weights, conical springs, or a more complex system synchronized with the respirator. This list is not exhaustive but it retraces the main

points that deserve to be raised. This is not insurmountable as divers have been using similar systems for decades.

AUTHOR CONTRIBUTIONS

NV wrote and the other authors proofread and validated the concepts and the manuscript. All authors contributed to the article and approved the submitted version.

FUNDING

This work should be attributed to the Institut de Recherches Biomédicales des Armées laboratories. It was supported by Grant No. SAN-1-0516 from the Direction Générale de l'Armement of the French army, Paris, France.

REFERENCES

- Arieli, R. (2008). The effect of over- or underfilling the soda lime canister on CO_2 absorption in two closed-circuit oxygen rebreathers. *Undersea Hyperbar. Med.* 35, 213–218.
- Gies, V., Barchasz, V., and Glotin, H. (2020). *Low Cost Artificial Ventilator Embedding Unsupervised Learning for Hardware Failure Detection*. La Garde: Smiot.
- Katz, I., Murdock, J., Palgen, M., and Farjot, G. (2017). A physiologically based model for denitrogenation kinetics. *Med. Gas Res.* 7, 256–259. doi: 10.4103/2045-9912.222449
- Logean, S. (2020). *Avec le Covid-19, le Monde Fait Face à Une Pénurie D'oxygène*. Lausanne: Le temps.
- Maves, R. C., Downar, J., Dichter, J. R., Hick, J. L., Devereaux, A., Geiling, J. A., et al. (2020). Triage of scarce critical care resources in COVID-19: an implementation guide for regional allocation an expert panel report of the Task force for mass critical care and the American college of chest physicians. *Chest.* 158, 212–225. doi: 10.1016/j.chest.2020.03.063
- Mycyk, M. B. (2020). Extracorporeal membrane oxygenation shows promise for treatment of poisoning some of the time: the challenge to do better by aiming higher. *Criti. Care Med.* 48, 1235–1237. doi: 10.1097/CCM.0000000000004431
- Navas-Blanco, J. R., and Dudaryk, R. (2020). Management of respiratory distress syndrome due to COVID-19 infection. *BMC Anesthesiol.* 20, 1–6. doi: 10.1186/s12871-020-01095-7
- See, H., and Slonim, A. D. (2020). Financial Toxicity after hospitalization; considerations in coronavirus disease 2019 recovery. *Criti. Care Med.* 48, 1233–1234. doi: 10.1097/CCM.0000000000004436
- Shaw, M., and Scott, D. (1998). Performance characteristics of a to and fro disposable soda lime canister. *Anaesthesia* 53, 454–460.
- Silva, D. S. (2020). Ventilators by lottery: the least unjust form of allocation in the COVID-19 Pandemic. *Chest* 158, 890–891. doi: 10.1016/j.chest.2020.04.049
- Silvanus, M., Mitchell, S. J., Pollock, N. W., Fränberg, O., Gennser, M., Lindén, J., et al. (2019). The performance of 'temperature stick' carbon dioxide absorbent monitors in diving rebreathers.

- Diving Hyperb. Med.* 49, 48–56. doi: 10.28920/dhm49.1.4 8-56
- Tobin, M. J. (2020). “Basing respiratory management of COVID-19 on physiological principles,” in *American Journal of Respiratory and Critical Care Medicine*, 201:1319–1320. doi: 10.1164/RCCM.202004-1076ED
- Wilmshurst, P. T. (2019). Immersion pulmonary oedema: a cardiological perspective. *Diving Hyperb. Med.* 49, 30–40. doi: 10.28920/dhm49.1.3 0-40
- Wingelaar, T. T., van Ooij, P. A. M., and van Hulst, R. A. (2017). Oxygen toxicity and special operations forces diving: hidden and dangerous. *Front. Psychol.* 8:1263. doi: 10.3389/fpsyg.2017.01263

Conflict of Interest: The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

Copyright © 2021 Vallée, Rives, Desruelle, Marzetti, Barchasz, Risso and Gies. This is an open-access article distributed under the terms of the Creative Commons Attribution License (CC BY). The use, distribution or reproduction in other forums is permitted, provided the original author(s) and the copyright owner(s) are credited and that the original publication in this journal is cited, in accordance with accepted academic practice. No use, distribution or reproduction is permitted which does not comply with these terms.